



**Certificate of calibration / Zertifikat der Eichung / Certificat d'étalonnage / Calibratie certificaat
Kalibreringscertifikat / Kalibrointsertifikaatti / Certificado de calibrado / Certificado de calibração**

**Automatic torque wrench / Automatischer Drehmomentschlüssel / Clé dynamométrique automatique / Automatische
momentsleutel / Momentnyckel / Automaattinen momenttlaavain / Llave dinamométrica / Chave dinamométrica**

117122

40-200 Nm

DIN EN ISO 6789 : 2003

(Norbar No.43218)

Accuracy / precision ($\pm 4\%/20^{\circ}\text{C}$)

Nm

Set torque	Min	Max	Actual readings			
40	38.4	41.6				
120	115.2	124.8				
200	192	208				

No. : _____

Operator : _____

Date : _____